We took the robust technology of the KALORIstar Arctic and spent it an outstanding design.

KALORIstar CT

The caloric examination is still one of the most popular methods in balance testing. The KALORIstar CT combines user friendly operation with a precise temperatured air flow.

Using air instead of water is more convenient for the patient. To achieve the same results as with water, the recommended temperatures for cold and warm irrigation are 24°C and 50°C.





BioMed Jena GmbH Am Egelsee 1 D-07743 Jena Germany


Phone: +49-3641 - 35690 -0 Fax: +49-3641 - 35690 -9 email: info@biomed-jena.de

WWW: www.biomed-jena.de



- » Ultra quite pump system
- » Temperature range: 20°C ... 51°C
- » Active cooling 7K below ambient temperature (@ 6l/min +/- 1l)
- » Temperatur accuracy: +/- 0,3K
- » Control Interface: Isolated USB
- » Automatic standby after irrigation
- » Length cable handle: 2m
- » Dimensions (WxHxD): 240x130x260 mm
- » Weight: 4 kg
- » Power consumption: max. 45W
- » Environmental: 10..40°C, <80% humidity
- » Standards: MDR Ann. VIII (EU 2017/745)

Thanks to the intelligent handle the customer only needs to send in the handle in case of service or maintanance. The KALORIstar ^{CT} can be fully remote controlled by the eVNG^{USB} software - the true one button operation in bithermal caloric test.



KALORIstar CT

Air **Irrigator** with active cooling technology (ACT)

01/2023 Rev02